SMS-FT100A-FN-02:2 0

Disposable probe cover

Infra-red ear thermometer

(Using for ear cannal measuring)

User's Instruction

Model: FT-100A (€ 0197

Probe membrane

Please read the instruction carefully before using.

The ear thermometer is a device capable of achieving infra-red temperature measurement when placed in the auditory canal of a subject. It is a safe means of measuring human body temperature through the ear.

Specification:

The standard we adopted is EN12470-5:2003 Clinical thermometers -- Part 5: Performance of infrared ear thermometers (with maximum device).

This device complies with the requirements of IEC 60601-1.IEC 60601-1-11 and IEC 60601-1-2.

Range of displayed temperature : 34.0 °C~44.0 °C (93.2 °F~111.2 °F) Display L°C(L°F) when the temperature is under 34.0°C(93.2°F) Display H°C(H°F) when the temperature is over 44.0°C(111.2°F)

Operating ambient:

Temperature: 16°C~35°C(60.8°F~95.0°F) Humidity: ≤80%RH

Storage and transport condition:

Temperature: -25°C to 55°C(-13°F to 131°F) Humidity: ≤85%RH

Display resolution temperature range: 0.1 °C (0.1°F)

Accuracy: $\pm 0.2^{\circ}$ C (from 35.5°C to 42.0°C) +0.4°F (from 95.9°F to 107.6°F)

Display: liquid crystal display, $3\frac{1}{2}$ digits

temperature value: display the maximum temperature

in measuring process temperature unit: centigrade or fahrenheit

display of memory: last ten memories

low voltage warning: the LCD display and then a beep sound is heard

Power consumption: ≤30mW

Battery: two 1.5V alkalescence batteries (AAA)

Battery life: 4000 takes

Dimension: 134mm by 37mm by 67mm

Net weight: Approx. 74g

Beeper sign: on/off, measuring finish and low voltage warning, etc.

Self-testing sequence: Press the 'on/recall' button to turn on the thermometer and all of symbol (See Fig A)

should be displayed on the LCD in one second.

Calibration frequency: two years, please done it by local Supervision Bureau.

Manufacturing and Calibration date: See the label in the batteries box (open the batteries cover use a Screwdriver).

Cautions

- 1. Before measuring, the thermometer shall be stabilized at least 30min under operating condition.
- 2. New Disposable probe cover required for each measurement, otherwise inaccurate readings will occur.
- 3. Please keep your ear canal clear, if not, the measuring result shall be inaccurate.
- 4. The symbol '?' on the LCD shows you can measure in the ear;
- 5. Please don't scratch the probe membrane, if not the thermometer shall lose the efficacy;
- 6. If the LCD no changes under pressing any button, please take the batteries out and put in again;
- 7. Please don't use the thermometer if your ear canal has been inflammation.
- 8. 1: Type BF equipment.
- 9. X : Product disposal instructions for electonic devices.
- The battery in this product complies with the requirements stated in European Directives 2006/66/EEC.

Note:

- In order to avoid spreading germs, please use a new disposable probe cover for each measurement.
- Check to ensure probe cover is fitted on firmly before use (please see the diagrams right).

If the probe cover is broken, discard and apply a new one immediately.




Fig A

How to use (℃ displays for example) Measuring human body temperature in the Ear Canal

- 1. Press the 'on/recall' button to turn on the thermometer, a beep sound is heard and the LCD displays as Fig A in one second. When the LCD displays as Fig B, now ready to measure.
- 2. Straighten the ear canal by pulling the outer ear up and back to give a clear view of the eardrum.
- For children under 1 year. Pull the ear straight back.
- For children ages 1 year to adult, Pull the ear up and back, (Caution: Many pediatricians suggest use of ear thermometers for children older than 6 months.)
- 3. While pulling the outer ear, insert the probe snugly into the ear canal (best to insert deep), press the 'scan' button, release it. When you hear a beep sound, the measurement is complete. Then remove the thermometer from the ear. The display show the measure temperature.
- 4. Measuring again: If measuring again, please unclinch the 'scan' button and wait for the ' 'symbol to display, then press 'scan' button to measure again. Advice: The time between each reading should be not less than 1min.
- 5. Memory seach: Press the 'on/recall' button, the last ten memories (NO.9-NO.0) will take turns to display on the LCD as Fig C and Fig D.
- 6. The LCD will displays 'Er 1' as Fig E and cannot displays temperature when the ambient temperature is out of the range of 10 °C to 35 °C.

Note: Before measuring, the thermometer shall be stabilized at the operating ambient condition for a minimum of 30 min.

7. Shut off: The thermometer shall automatically shut off without manipulating in one minute.





Changing from Celsius to Fahrenheit (only for switchable type)

To switch the display between C and F. turn the unit on. Press and hold the on/recall button for 3 seconds, the display will show as Fig. F. the unit will prepare for a measurement.



Product disposal

Please ensure environmental protection. Batteries do not belong in the domestic waste. Please hand them in at collection point or the municipal recycle material centre as special waste.

This symbol on products and/or accompanying documents means that consumed electronic products must not be mixed with conventional domestic waste. Take these products to the corresponding collection points for correct treatment and revoling, where they will be accepted free of charge. For more information on the closest collection point, Please enquire with your local authorities.

Cleaning and Storage

- 1. Store thermometer and the **Disposable probe cover** in a dry location free from dust and contamination and away from direct sunlight. The ambient temperature at the storage location should remain fairly constant and within the range of -25°C to 55°C.
- 2. Use an alcohol swab or cotton swab moistened with alcohol (70% Isopropy) to clean the thermometer casing and the measuring probe. Ensure that no liquid enters the interior of the thermometer
- 3. Never use abrasive cleaning agents, thinners or gasoline for cleaning and never immerse the instrument in water or other cleaning liquids. Take care not to scratch the surface of the probe membrance or display

Replacing the batteries

- 1. When voltage of the batteries are low, the LCD will display '====' symbol, please replace two new batteries in unit. Thethermometer cannot work accurately under the condition of low voltage.
- 2. The thermometer is supplied with two 1.5V ALKALESCENCE BATTERIES (AAA). Insert new batteries when the low voltage symbol appears on the LCD.
- 3. Screw off the bolt and remove the battery cover and take out the old batteries.
- 4. Place two new batteries according to the '+' or '-'
- 5. The disposal of the probe cover, battery and device shall comply with the local environment requirements.
- 6. Please take out the batteries to avoid battery leaking if unit not used for over six months.

Quality warranty

This thermometer is guaranteed for one year under proper use according the operating manual from the date that the product is sent out the factory.





